

## **Purchasing Division**

## **ADDENDUM NO. 1**

DATE: November 22, 2023

FROM: City of Grand Junction Purchasing Division

TO: All Offerors

RE: Obsolete Auto Titrator and Auto Sampler IFB-5332-23-KH

Offerors responding to the above referenced solicitation are hereby instructed that the requirements have been clarified, modified, superseded and supplemented as to this date as hereinafter described.

Please make note of the following clarifications:

1. **Q**: (Specification 19 - Must have minimum of 3-year warranty) We would like to know which of the 3 below options are to be included in this. We ask since vendors have specific exclusions related to multi-year warranty's such as those listed in Option 1: Option 1: 3-year warranty - Warranty service performed after the 1st year excludes travel and labor charges. Travel and labor will be charged at published rates for any warranty repairs performed after the 1st year. - Warranty only remains valid on the condition that an annual preventive maintenance service is provided by a MANTECH factory trained representative. Option 2: 3-year warranty - Warranty service includes parts, labour and travel expenses for 3 years. Yearly on-site preventative maintenance included at 12-14 months and 24-26 months. Option 3: 3-year warranty - Warranty service includes parts, labour and travel expenses for 3 years.

A: Option 2.

- 2. **Q:** (Specification 7 Must be able to support at minimum two titrating burettes) We understand the capability to "support" a minimum of two, however, is one (1) or two (2) titrating burettes to be included in the proposal?
  - **A:** Two titrating burettes should be included in the proposal.
- 3. **Q:** (Specification 11 Must be capable of running EPA approved methods for alkalinity, volatile acids, and sample neutralization) Please provide the corresponding EPA approved method that are being referenced, for example, Standard Methods, ASTM, EPA, etc. For example, Alkalinity: Standard Methods 2320 B
  - A: Method specific would be Standard Methods 2320 B.

The original solicitation for the project noted above is amended as noted.

All other conditions of subject remain the same.

Respectfully,

Kassy Hackett, Buyer

City of Grand Junction, Colorado